

### Certificate of Analysis

#### Laboratory Reference:230124-027

<b>Attention:</b>	Nz Queenstown Lab Results	<b>Final Report:</b>	<b>496645-0</b>	<b>Replaces Report</b>	<b>494816-0</b>
<b>Client:</b>	<b>VEOLIA WATER</b>	<b>Report Issue Date:</b>	<b>16-Feb-2023</b>		
<b>Address:</b>	<b>74 Glenda Drive, Frankton, Queenstown, 9300</b>	<b>Received Date:</b>	<b>24-Jan-2023</b>		
		<b>Sampled By:</b>	<b>HB</b>		
<b>Client Reference:</b>		<b>Laboratory Activity Dates:</b>	<b>24-Jan-2023</b>	<b>-</b>	<b>02-Feb-2023</b>
<b>Purchase Order:</b>	<b>7300228298</b>	<b>Quote Reference :</b>	<b>48</b>		

Amended report: Please note added metals results for samples 230124-027-1 (Western Source) and 230124-027-2 (Beacon Point Source).

#### Sample Details

	WATERS	WATERS	WATERS
<b>Lab Sample ID:</b>	<b>230124-027-1</b>	<b>230124-027-2</b>	<b>230124-027-3</b>
<b>Client Sample ID:</b>	S2905	S2906	WAN002WA
<b>Sample Date/Time</b>	24/01/2023 10:00	24/01/2023 10:30	24/01/2023 10:15
<b>Description:</b>	Western Source	Beacon Point Source	WAN002WA 213 Brownston Street - Wanaka

#### Micro Summary View

	MPN/100mL			
Escherichia coli		<1.0	<1.0	-
Total Coliforms		30	290	-

#### Chemistry Detailed

##### Anions

	mg/L			
Chloride		0.643 ± 0.045	0.657 ± 0.046	-
Nitrate (as N)		0.018 ± 0.0023	0.022 ± 0.0028	-
Sulphate		6.69 ± 0.47	3.7 ± 0.44	-

#### Sample Parameters and Field Testing

	mg/L			
Residual Free Chlorine (as Cl <sub>2</sub> )		-	-	0.73 ± 0.037
Sampled by Watercare Laboratory Services		Yes	Yes	Yes

#### General Testing

	mg/L			
Total Alkalinity (as CaCO <sub>3</sub> )		30	30	-
True Colour (Dissolved)	Hazen Units	<5.0 *	<5.0 *	-

#### Metals

##### Total Metals by ICP-MS—Trace (Default Digest)

	mg/L			
Antimony (Total)		<0.001	<0.001	-
Arsenic (Total)		0.00068 ± 0.00014	0.0006 ± 0.00012	-
Barium (Total)		0.0041 ± 0.00068	0.0046 ± 0.00076	-
Cadmium (Total)		<0.00005	<0.00005	-
Calcium (Total)		11 ± 1.8	11 ± 1.7	-
Chromium (Total)		<0.0005	<0.0005	-
Copper (Total)		0.0034 ± 0.00063	0.0069 ± 0.0013	-
Iron (Total)		0.012 ± 0.0025	0.0086 ± 0.0017	-
Lead (Total)		0.0012 ± 0.00021	0.0014 ± 0.00024	-
Magnesium (Total)		0.6 ± 0.13	0.59 ± 0.13	-
Manganese (Total)		0.0014 ± 0.00023	0.001 ± 0.00017	-
Mercury (Total)		<0.00005	<0.00005	-
Nickel (Total)		0.00021 ± 0.000038	0.00042 ± 0.000074	-
Sodium (Total)		0.83 ± 0.17	0.81 ± 0.17	-

Results marked with \* are not accredited to International Accreditation New Zealand. A dash indicates no test performed.

Where samples have been supplied by the client, they are tested as received.

The results of analysis contained in this report relate only to the sample(s) tested. Where sample collection was performed by the laboratory, the results of analysis contained in this report relate only to the sample(s) collected.



#### Queenstown

74 Glenda Drive, Frankton, Queenstown, 9300  
 T: (03) 409 0559

#### Invercargill

142 Esk Street, Invercargill, 9810  
 T: (03) 214 4040

## Analyst's Notes

The Nitrate (as N) analysis for sample Western Source commenced beyond the holding time of 2 Days

The Nitrate (as N) analysis for sample Beacon Point Source commenced beyond the holding time of 2 Days

## Reference Methods

The sample(s) referred to in this report were analysed by the following method(s)

Analyte	Method Reference	MDL	Samples	Location
<b>Micro Summary View</b>				
Escherichia coli	APHA (online edition) 9223 B	1 MPN/100mL	1, 2	Queenstown
Total Coliforms	APHA (online edition) 9223 B	1 MPN/100mL	1, 2	Queenstown
<b>Chemistry Detailed</b>				
<b>Anions</b>				
Chloride	APHA (online edition) 4110 B	0.02 mg/L	1, 2	Auckland
Nitrate (as N)	APHA (online edition) 4110 B	0.002 mg/L	1, 2	Auckland
Sulphate	APHA (online edition) 4110 B	0.02 mg/L	1, 2	Auckland
<b>Sample Parameters and Field Testing</b>				
Residual Free Chlorine (as Cl <sub>2</sub> )	APHA (online edition) 4500-Cl G	0.02 mg/L	3	At site
Sampled by Watercare Laboratory Services	AS/NZS 5667 Part 4 (Lakes), Part 5 (Drinking Water), Part 6 (Rivers/Streams), Part 7 (Boilers), Part 9 (Marine), Part 10 (Waste Waters), Part 11 (Groundwaters) and In-House Methods		All	At site
<b>General Testing</b>				
Total Alkalinity (as CaCO <sub>3</sub> ) by Titration	APHA (online edition) 2320 B	1 mg/L	1, 2	Auckland
True Colour (Dissolved) by Colour Comparator	APHA (online edition) 2120 B	5 Hazen Units	1, 2	Auckland
<b>Metals</b>				
<b>Total Metals by ICP-MS—Trace (Default Digest)</b>				
Antimony (Total)	APHA (online edition) 3125 B by ICPMS	0.001 mg/L	1, 2	Auckland
Arsenic (Total)	APHA (online edition) 3125 B by ICPMS	0.0001 mg/L	1, 2	Auckland
Barium (Total)	APHA (online edition) 3125 B by ICPMS	0.0002 mg/L	1, 2	Auckland
Cadmium (Total)	APHA (online edition) 3125 B by ICPMS	0.00005 mg/L	1, 2	Auckland
Calcium (Total)	APHA (online edition) 3125 B by ICPMS	0.01 mg/L	1, 2	Auckland
Chromium (Total)	APHA (online edition) 3125 B by ICPMS	0.0005 mg/L	1, 2	Auckland
Copper (Total)	APHA (online edition) 3125 B by ICPMS	0.0002 mg/L	1, 2	Auckland
Iron (Total)	APHA (online edition) 3125 B by ICPMS	0.002 mg/L	1, 2	Auckland
Lead (Total)	APHA (online edition) 3125 B by ICPMS	0.0001 mg/L	1, 2	Auckland
Magnesium (Total)	APHA (online edition) 3125 B by ICPMS	0.001 mg/L	1, 2	Auckland
Manganese (Total)	APHA (online edition) 3125 B by ICPMS	0.0005 mg/L	1, 2	Auckland
Mercury (Total)	APHA (online edition) 3125 B by ICPMS	0.00005 mg/L	1, 2	Auckland
Nickel (Total)	APHA (online edition) 3125 B by ICPMS	0.0001 mg/L	1, 2	Auckland
Sodium (Total)	APHA (online edition) 3125 B by ICPMS	0.1 mg/L	1, 2	Auckland
<b>Preparations</b>				
Digest for Total Metals in Liquids	APHA 3030E Modified ( 4:1 Nitric:Hydrochloric Acid: 95°C 2 hours)		1, 2	Auckland
Membrane Filtration (0.45 µm)	APHA (online edition) 4500-P B (preliminary filtration)		1, 2	Auckland
<i>The method detection limit (MDL) listed is the limit attainable in a relatively clean matrix. If dilutions are required for analysis the detection limit may be higher. For more information please contact the Compliance and Projects Manager.</i>				

Samples, with suitable preservation and stability of analytes, will be held by the laboratory for a period of two weeks after results have been reported, unless otherwise advised by the submitter.

Watercare Laboratory Services is a division of Watercare Services Limited.

**This report may not be reproduced, except in full, without the written authority of the Compliance and Projects Manager.**



### Queenstown

74 Glenda Drive, Frankton, Queenstown, 9300  
T: (03) 409 0559

### Invercargill

142 Esk Street, Invercargill, 9810  
T: (03) 214 4040



Chandra Sharma  
KTP Signatory



Anel Du Preez  
KTP Signatory



David Hoekendijk  
KTP Signatory



**Queenstown**  
74 Glenda Drive, Frankton, Queenstown, 9300  
T: (03) 409 0559

**Invercargill**  
142 Esk Street, Invercargill, 9810  
T: (03) 214 4040